

Certification of EO GAS Sterilization

Date: 5-02-2022

Customer	TCM Supply Corporation DBA: WorldWide Tattoo Supply										
Product codes	Product name	Lot No									
/											
Sterilization mode	Chamber NO.	Cartridge type									
EO-Gas	2#	0.25; 0.30; 0.35mm		0.25, 0.30, 0.35mm : RSB; RLB; M1B; M1TB; M1TB-T							
Sterilization parameters		Validated	Tolerance	Actual							
Preconditioning temperature (°C)		50	±4	50.5-53.5							
Preconditioning RH (RH %)		55	±10	49-55							
Preconditioning Time (min)		360	±20	350							
Initial Evacuation Vacuum (kpa)		-25	±2	-25							
Keep Pressure (min)		20	±10	15							
Weight of EO Injection (kg/25m3)		15	±1.25	15							
Rate of EO Injection (kpa/min)		0.5	±0.2	0.5							
EO-Concentration (mg/L)		600	±50	600							
EO Injection Pressure Changed (kpa)		>25	//	S:-25.0E: 8.0							
Exposure Time (min)		720	±20	720							
Sterilization Temperature (°C)		52	±3	50.5-53.5							
Chamber RH of Sterilization (%RH)		55	±10	49.5-55.5							
Post Evacuation Vacuum (kpa)		-20	±2	-20.5							
The Number of ventilator (times)		6	//	6							
Aeration Temperature (°C)		42	±5	38.5-40.5							
Aeration Time (day)		7	//	7							
Strain	Bocillus Ptrophoeus ATCC9372	Nominal Population (cfu/unit)				1.7x10 ⁶					
Lot NO.	200825	D-Value (min)				3					
Culture temp.	36°C	Culture time (days)				7					
Culture date	2022 - 2025										
Culture time	1	2	3	4	5	6					
Observed item	-	-	-	-	-	-					
Positive control	+	+	+	+	+	+					
BI test of sterility		Conformity									
Conclusion:											
The sterilization cycles was performed as per the validated parameters with no deviations.											
The sterile Assurance Level (SAL) of 10 ⁶ is assured											

Approved by: *Jad L.K.*

Company stamp:

